



## Akers Biosciences, Inc.

Heparin/PF4 Antibody Serum Panel

Lot # 22080

Expiry 2011/11

Certificate of Quality Data Sheet

<u>Panel ID#</u>	<u>Matrix</u>	<u>GTI<sup>®</sup> PF4 ENHANCED<sup>®*</sup> ELISA Assay</u>	<u>PIFA<sup>®**</sup> H/PF4 Antibody Determination</u>
21775	Serum	0.492 (Positive)	Positive
21963	Serum	0.213 (Negative)	Negative

The data provided within is offered for informational purposes. Akers Biosciences, Inc. does not claim that others can duplicate these test results exactly. According to the GTI<sup>®</sup> PF4 ENHANCED<sup>®</sup> ELISA Assay Package Insert, O.D. readings obtained from duplicate tests should fall within  $\pm 20\%$  of the mean of the two values. All testing was performed at Akers Biosciences, Inc. by individuals who routinely use these procedures.

\* GTI<sup>®</sup> and PF4 Enhanced<sup>®</sup> are registered trademarks of Gen-Probe GTI Diagnostics, Inc., Waukesha, Wisconsin USA

\*\* PIFA<sup>®</sup> is a registered trademark of Akers Biosciences, Inc., Thorofare, New Jersey USA